










Hose Order Guide

Z - A B - C

Length in Feet:

Gas:	Inlet Connections:	Outlet Connections:
O2 = Oxygen	F = Female DISS	F = Female DISS
AIR = Medical Air	FH = Female Hand-Tight DISS	FH = Female Hand-Tight DISS
VAC = Medical Vacuum	M = Male DISS Check Valve	M = Male DISS Check Valve
N2O = Nitrous Oxide	MB = Male DISS Body Adaptor	MB = Male DISS Body Adaptor
N2 = Nitrogen	PF = Puritan Female	PF = Puritan Female
CO2 = Carbon Dioxide	PM = Puritan Male	PM = Puritan Male
AGSS = Anaesthetic Gas Scavenging System	OF = Ohio Female	OF = Ohio Female
WAGD = Waste Anesthetic Gas Disposal	OM = Ohio Male	OM = Ohio Male
EVAC = Laser Smoke Evacuation	XF = Oxequip Female	XF = Oxequip Female
IA = Instrument Air	XM = Oxequip Male	XM = Oxequip Male
	MF = Medstar Female	MF = Medstar Female
	MM = Medstar Male	MM = Medstar Male
	CF = Chemtron Female	CF = Chemtron Female
	CM = Chemtron Male	CM = Chemtron Male
	SF = Schrader Female	SF = Schrader Female
	SM = Schrader Male	SM = Schrader Male
	F75 = F 3/4" NPT	F75 = F 3/4" NPT
	M75S = M 3/4" NPT Swivel	M75S = M 3/4" NPT Swivel

ISO Hoses		
O2 (CSA)		White, 1/4" ID
VAC (CSA)		Yellow, 5/16" ID
AIR (CSA)		Black & White, 1/4" ID
N2O (CSA)		Blue, 1/4" ID
N2 (CSA)		Black 1/4" ID
CO2 (CSA)		Gray, 1/4" ID
AGSS (CSA)		Magenta, 1/4" ID (High Vac Style)
EVAC (CSA)		Magenta, 1/4" ID (High Vac Style)
IA (CSA)		Black 1/4" ID

Example- AIR-FHM-3

